

"Express Mail" Mailing Label Number EL545796248US  
Date of Deposit: February 13, 2001  
I hereby certify that this paper or fee is being deposited  
with the United States Postal Service "Express Mail  
Post Office to Addressee" service under 37 C.F.R. 1.10 on the  
date indicated above and is addressed to:  
Assistant Commissioner for Patents, Washington, D.C. 20231.

*Keith A. Johnson*  
By: Keith JOHNSON

## LIMITED-ANGLE FREQUENCY-DISTANCE RESOLUTION RECOVERY IN NUCLEAR MEDICINE IMAGING

### Background of the Invention

5 The present invention relates to the diagnostic  
imaging art. It finds particular application in  
conjunction with nuclear single photon emission computed  
tomography (SPECT) medical imaging and will be described  
with particular reference thereto. However, the invention  
will also find application in conjunction with other types  
of non-invasive diagnostic imaging.

10 Heretofore, diagnostic images have been  
generated from single and multiple-head nuclear cameras.  
Typically, a patient positioned in an examination region  
is injected with a radiopharmaceutical. Heads of the  
nuclear camera are positioned closely adjacent to the  
15 patient to monitor the radiopharmaceutical. Typically,  
the heads are stepped in increments of a few degrees  
around the patient until 360° of data have been acquired.  
That is, projection data along directions spanning 360°  
are collected. With multiple-head systems, projections  
20 along each direction need only be collected with one of  
the heads and be assembled into a complete data set.

Each detector head carries a collimator which  
defines a path along which it can receive radiation.  
However, due to the finite length and dimensions of the  
collimator, each incremental area of the detector head  
25 actually views an expanding cone. Thus, with increasing  
depth into the patient away from the detector head, the  
region from which a sensed radiation event originated  
expands. This creates depth-dependent blurring and

uncertainty in the resultant image data. This error is a non-stationary convolution which is difficult to deconvolve. However, when the patient is viewed over a full 360°, the angular data sets are periodic in 2π radians. By transforming the full data sets into the frequency domain with a Fourier transform fit to the sampling intervals, the non-stationary deconvolution is reduced to a stationary deconvolution problem, particularly for high frequencies.

Although these prior art resolution recovery techniques work well on full data sets, cardiac imaging is typically done using an incomplete data set. More specifically, in a three-head camera system where only two heads collect the emission data while the third head is used to collect transmission data, the gantry is rotated by about 102° to generate the equivalent of only about 204° of emission data. The prior art resolution recovery techniques do not work on partial data sets whose data is not periodic in 2π radians.

The present invention contemplates a new and improved method and apparatus which overcomes the above-referenced problems and others.

#### Summary of the Invention

In accordance with one aspect of the present invention, a method of diagnostic imaging is disclosed. A plurality of projection data sets are collected at each of a plurality of angles around a subject. The projection images are collected over less than 360°. A resolution recovery process is performed on the projection data sets. The resolution recovered projection data sets are reconstructed into an image representation.

In accordance with another aspect of the present invention, a method of diagnostic imaging is disclosed. A gantry moves a detector head in a continuous angular orbit about a subject in an examination region. Data is collected during the continuous orbit and sorted into a

plurality of projection data sets corresponding to each of a plurality of angular increments around a subject. A resolution recovery process is performed on the projection data sets. The resolution recovered projection data sets  
5 are reconstructed into an image representation.

In accordance with yet another aspect of the present invention, a diagnostic imaging apparatus is disclosed. At least one detector head detects incident radiation. A collimator mounted to the detector head  
10 limits trajectories along which radiation is receivable by the head. A movable gantry moves the detector head around a subject in an examination region. A data acquisition system acquires projection data sets from the detector head at angular increments spanning less than  $360^\circ$ . A  
15 zero-filling processor generates zero-filled data sets between the actually collected projection data sets, to create  $360^\circ$  of data sets. A smoothing processor smooths interfaces between the actually collected and zero-filled data sets. A resolution recovery processor operates on  
20 the smoothed data sets. A reconstruction processor reconstructs the resolution recovered data sets into a three-dimensional image representation. An image memory stores the three-dimensional image representation.

One advantage of the present invention is that  
25 it accurately restores limited-angle data sets.

Another advantage of the present invention is that it restores continuously scanned data sets.

Another advantage of the present invention is that it processes image data in a clinically feasible  
30 time.

Still further advantages and benefits of the present invention will become apparent to those of ordinary skill in the art upon reading and understanding the following detailed description of the preferred  
35 embodiments.

### Brief Description of the Drawings

The invention may take form in various components and arrangements of components, and in various steps and arrangements of steps. The drawings are only  
5 for purposes of illustrating preferred embodiments and are not to be construed as limiting the invention.

FIGURE 1 is a diagrammatic illustration of a nuclear medicine imaging system in accordance with the present invention;

10 FIGURE 2 is a diagrammatic illustration of the formation of the modified data set which is periodic in  $2\pi$  radians, based on the non-periodic limited-angle data set;

FIGURE 3 shows the coordinate system used in data acquisition and image reconstruction.

### Detailed Description of the Preferred Embodiments

With reference to FIGURE 1, a nuclear camera system 10 includes a plurality of detector heads 12, in the preferred embodiment three detector heads 12<sub>1</sub>, 12<sub>2</sub>, and 12<sub>3</sub>. Two of the three heads are typically used to obtain  
20 emission data, while the third head is used to obtain transmission data. Of course, other numbers of detector heads can also be utilized. Each of the detector heads includes a collimator 14<sub>1</sub>, 14<sub>2</sub>, and 14<sub>3</sub>. In the preferred embodiment, the collimators collimate incoming radiation  
25 from a subject 16 to parallel rays. However, because the collimators have finite size, each collimator permits rays which lie along a corresponding cone to pass to the detector head. The cone expands with depth into the patient from the detector head.

30 Each of the detector heads includes a reconstruction system 20<sub>1</sub>, 20<sub>2</sub>, and 20<sub>3</sub>, which determines the coordinates on a face of the detector head in the longitudinal or z-direction of the patient and the transverse direction across the detector head. With the  
35 detector head in a single orientation, scintillation events are collected for a preselected period of time to

generate a projection image representation. After the preselected data acquisition duration, a rotating gantry 22 rotates all three detector heads concurrently a short angular distance, e.g.,  $3^\circ$ . In the new location, each of the detector heads collects another projection image. An angular orientation monitor 24 determines the angular orientation of each of the heads at each angular data collection position.

A data acquisition system 30 receives each of the projection images and an indication of the angle along which it had been taken. This data is stored in a three-dimensional memory 32 in longitudinal (z), lateral (s), and angular ( $\phi$ ) coordinates.

With continuing reference to FIGURE 1 and further reference to FIGURE 2, the rotating gantry 22 rotates the detector heads over  $102^\circ$  for cardiac imaging. Rotation over  $102^\circ$  generates the equivalent of about  $204^\circ$  of emission data 110 collected by two camera heads. Each data set is collected at equal angular increments, e.g.,  $3^\circ$ . FIGURE 2 shows an incomplete data set with no data collected between  $204^\circ$  and  $360^\circ$  in the angular dimension. Simultaneously, the third camera collects transmission data over the  $102^\circ$  gantry rotation.

With continuing reference to FIGURE 2 and reference again to FIGURE 1, a resolution recovery system 42 includes an angular displacement dimension resolution recovery system and optionally includes resolution enhancement sub-system 46 for the axial dimension. The resolution recovery system in the angular direction ( $\phi$ ) has a zero-filling processor 50 which creates zero magnitude projection data sets 112 at each of the  $3^\circ$  intervals between  $204^\circ$  and  $360^\circ$ . In this manner, a function which is periodic in  $2\pi$  radians is created. In order to prevent Gibbs' ringing, the sudden discontinuities at each end of the actually collected data 110 between the actually collected and zero-filled data 112 are smoothed by a smoothing processor 56. In the

preferred embodiment, the magnitude of the end points are each cut in half producing modified data points 114 and 116. Optionally, other smoothing functions which span several points at each end are also contemplated.

5           A Fourier transform processor 60 transforms the data into the frequency domain. The Fourier transform is selected to match all of the sampling points including the zero-filled points in the angular dimension. A stationary deconvolution processor 62 operates on the  
10 frequency-spaced data. In frequency space, the deconvolution problem reduces to a stationary deconvolution problem, particularly for high frequencies. Because high frequencies correspond to fine detail, it is the high frequencies which correspond to the resolution to  
15 be optimized. The additional deconvolution 46 may optionally also be performed to improve image resolution in the z-direction. After the data has been deconvolved, an inverse Fourier transform processor 64 transforms the data from the frequency domain back into real space.  
20 Optionally, a three-dimensional memory 66 stores the resolution recovered sets of projection data. A reconstruction processor 68 reconstructs the projection data sets, using filtered back-projection, iterative reconstruction, or other techniques which are well-known  
25 in the art, and stores the resultant image in a three-dimensional image memory 70. A video processor 72 withdraws selected portions of the reconstructed image and converts it to appropriate format for display on a human-readable monitor 74, such as a video monitor, CCD  
30 display, active matrix, or the like. The video processor may withdraw selected slices, three-dimensional renderings, projection views, or the like.

          In one alternate embodiment, the movable gantry  
22 rotates the detector heads continuously. Although the  
35 detector heads rotate continuously, the data acquisition system 30 bins the collected data into regular angular intervals, e.g., 3°. Because the data is collected over

3° of rotation, there is an additional blurring component. As described in greater detail below, the stationary deconvolution processor 62 deconvolves the Fourier space data with respect to the blurring caused by the continuous motion and treats data collected over a few degrees as if it were all collected at precisely the same angle.

With reference to FIGURE 3, and looking now to details of the resolution recovery system, the working principles are explained in two-dimensions and circular orbits, but the generalization to three-dimensions is straightforward and generalization to non-circular orbits is known to those conversed in the art. The third dimension, z, is perpendicular to the plane defined by the x-y axes. The scanner collects data in coordinates (s,φ) which are the sinogram coordinates of the Radon transform coordinates. The unblurred or undegraded Radon transform of the object o is the line integral along t, the axis perpendicular to s. In a nuclear medicine tomographic device, the object o is blurred by a point response function g. This process is modeled by a convolution in s. The amount of blurring depends on the depth t. Hence, the blurring function is a non-stationary convolution of o with g. The result is the blurred Radon transform p. For limited-angle tomography, this operation of the scanner is represented by:

$$p(s, \phi) = \int_{-\infty}^{\infty} \int_{-\infty}^{\infty} g(s-s', t) o(s'\theta + t\theta^\perp) ds' dt \quad (1),$$

where  $\theta$  and  $\theta^\perp$  are two-dimensional unit vectors aligned with the (s,t) axes:

$$\theta: [\cos\phi, \sin\phi] \quad \text{and} \quad \theta^\perp: [-\sin\phi, \cos\phi] \quad (2).$$

A point source located at x in the object will have the point-source projection:

$$p_{\delta} = \int_{-\infty}^{\infty} \int_{-\infty}^{\infty} g(s-s', t) |J| \delta(s'-\mathbf{x} \cdot \boldsymbol{\theta}, t-\mathbf{x} \cdot \boldsymbol{\theta}^{\perp}) ds' dt = g(s-\mathbf{x} \cdot \boldsymbol{\theta}, \mathbf{x} \cdot \boldsymbol{\theta}^{\perp}) \quad (3),$$

where  $|J|$  is the Jacobian of the change of coordinates from  $(x, y)$  to  $(s, t)$  and is equal to one because it is a rotation. The two-dimensional Fourier transform of Equation (3) gives:

$$\begin{aligned} \hat{p}_{\delta}(\omega, n) &= \int_0^{2\pi} e^{-in\phi} \int_{-\infty}^{\infty} e^{-i\omega s} g(s-\mathbf{x} \cdot \boldsymbol{\theta}, \mathbf{x} \cdot \boldsymbol{\theta}^{\perp}) ds d\phi \\ &= \int_0^{2\pi} e^{-in\phi} e^{-i\omega \mathbf{x} \cdot \boldsymbol{\theta}} \hat{g}(\omega, \mathbf{x} \cdot \boldsymbol{\theta}^{\perp}) d\phi \approx \hat{g}\left(\omega, \mathbf{x} \cdot \boldsymbol{\theta}^{\perp} = \frac{n}{\omega}\right) \end{aligned} \quad (4).$$

- 5 The  $\phi$ -integral is evaluated using the principle of stationary phase, an approximation process familiar to those knowledgeable in the art, which asymptotically converges to the exact value of the  $\hat{p}_{\delta}$  integral at high frequencies. In Equation (4), it will be noted that the  
10 solution does not depend on the location of the point source. Hence, it can be applied to any point source or a weighted collection of point sources.

Furthermore, if the scanning arc is limited to something less than  $360^\circ$  or  $2\pi$  radians, the basic result  
15 of stationary phase, that the distance  $t$  is given by:

$$t = \mathbf{x} \cdot \boldsymbol{\theta}^{\perp} = \frac{n}{\omega} \quad (5)$$

is unchanged, because every distance is represented in each planar projection image. This enables the limits of integration in Equation (4) to be replaced with an arc or a series of disjoint arcs and Equation (5) still holds to  
20 good approximation.

In operation, the digitized three-dimensional data set  $p(s_j, z_k, \phi_m)$  is collected. The digitization assumes equal increments in each coordinate sample space. For example, the angular increment  $m$  is taken at intervals  
 5  $\phi_m = 2\pi m/N$  radians, where  $N$  is an integer.

The depth dependent, point response function  $g(s, z, t)$  is obtained. When the scan is over a limited angle arc described by the interval  $[\phi_u, \phi_v]$ , a digital input data set is formed according to the following  
 10 process:

$$p_{in}(s_j, z_k, \phi_w) = \begin{cases} 0, & \text{for } 0 \leq W < \min(u, v) \\ \frac{1}{2}p(s_j, z_k, \phi_u), & \text{for } W = u \\ p(s_j, z_k, \phi_w) & \text{for } u < W < v \\ \frac{1}{2}p(s_j, z_k, \phi_v), & \text{for } W = v \\ 0, & \text{for } \max(u, v) < W < N \end{cases} \quad (6).$$

The endpoint projection frames are multiplied by one-half to prevent excessive ringing or Gibbs' phenomena. Other smoothing functions for smoothing the sudden data discontinuity are also contemplated. The Fourier  
 15 transform with respect to angle is selected to have dimension  $N$ , no more and no less, to preserve the cyclic nature of the data. The three-dimensional Fourier transform applied by the Fourier transform processor 60 of the modified projection data  $p_{in}$  is denoted by:

$$\hat{p}_{in}(\omega_s, \omega_z, n) \quad (7).$$

20 The filter function:

$$\hat{g}\left(\omega_s, \omega_z, \frac{n}{\omega_s}\right) \quad (8)$$

is defined. The filter function is used in a regularized inverse filter such as a noise reduction filter or a Metz filter to perform the stationary deconvolution 62. The

inverse Fourier transform processor 64 takes the inverse Fourier transform of the filtered image to obtain projection data that possesses improved resolution.

In the continuous scanning embodiment, the camera moves according to a preprogrammed orbit. In the preferred embodiment, the three heads move simultaneously in a complex patient dependent orbit. Continuous motion is advantageous for reducing motion complexity. However, continuous scanning degrades resolution. Preferably, the stationary deconvolution module includes a component to compensate for this degraded resolution.

Continuous scan projection data  $p_c$  is related to the step-and-shoot projection data  $p$  by:

$$p_c(s, z, \phi) = \frac{1}{\Delta\phi} \int_{-\Delta\phi/2}^{\Delta\phi/2} p(s, z, \phi') d\phi' \quad (9).$$

The Fourier transform relationship is:

$$\hat{p}_c(\omega_s, \omega_z, n) = \hat{p}(\omega_s, \omega_z, n) \frac{\sin(n\Delta\phi/2)}{n\Delta\phi/2}, -\frac{N}{2} + 1 \leq n \leq \frac{N}{2} \quad (10).$$

These relationships hold true for non-circular and limited-angle orbits as well. The filter function for continuous sampling  $\hat{g}_c$ , is modified to include the weighted sine function from Equation (10):

$$\hat{g}_c\left(\omega_s, \omega_z, \frac{n}{\omega_s}\right) = \frac{\sin(n\Delta\phi/2)}{n\Delta\phi/2} \hat{g}\left(\omega_s, \omega_z, \frac{n}{\omega_s}\right) \quad (11).$$

In accordance with another alternate embodiment, in the step-and-shoot imaging mode, the actually collected data integrals may be disjoint. It is not necessary for all of the angular views to be adjacent as in the example above.

The invention has been described with reference to the preferred embodiments. Obviously, modifications

[illegible]